



ALL DIMENSIONS ARE IN mm.

NOTE :- THE MIN. DISTANCE OF 300mm FROM THE WALL SHOULD BE KEPT FOR THE FOUNDATION BLOCK SIDES.

FOUNDATION BOLT DETAILS AS PER IS:5624

DATE : 27.06.09
 DRN. : SSP
 CHD. :
 APPD. :

DRG. No.: MTT - E - 200Nm - FP - 01

FOUNDATION PLAN FOR TORSION TESTING MACHINE (MODEL MTT-E-100/200 Nm)

META TEST INSTRUMENT PVT. LTD. B-11 M.I.D.C. MIRAJ