

**FRI**

Prosperity through research

# Fire Research Institute

22/44, Ganganagar P. O., ICHALKARANJI - 416 116.  
(Dist.Kolhapur) Maharashtra State, INDIA.

■ Phone : (0230) 2441475 ■ Fax : (0230) 2440191

■ E-mail : fri@dataone.in ■ Website : www.fiegrouppindia.com

**Accreditation No. NABL C0066**

## CALIBRATION CERTIFICATE OF FORCE MEASURING DEVICE

Date of calibration : 12/01/2014  
Next calibration due on : 12/03/2016Certificate No : FRI/01/14/6605  
Page No. : 1/2

Calibrated for Krutam Techno Solutions Pvt. Ltd.  
64/B, G.I.D.C. Makarpura Industrial Estate,  
Behind Fire Brigade Station,  
Vadodara - 390 010

Customer Reference No. D.C.No. 01 Dated on 10/01/2014

Details of instrument under calibration

Instrument	: Dynamometer
Id No.	: 104 (KTPL/F/003)
Capacity	: 50 kN
Type	: Bow
Dial Gauge Make	: Baker
1 Div	: 0.01 mm
Dial Gauge No.	: VCI 655
Fulcrum Pin No.	: 104

Date of receipt 11/01/2014

Mode of calibration Compression

Machine used for calibration Dead Weight Force Machine(FIE-DWP-003)

Traceability NPL,Cert.No.12031657/D.5.05/C-084 valid upto 07/07/14




Dial gauge setting Large pointer at 12 O' Clock position  
Small pointer at 2 revolutions

Temperature 24°C

Correction of temperature variation Apply  $\pm 0.027\%$  correction to each reading  
for each °C rise or fall of temperature.

**Note**

- 1) Compression test were made out by using compression pads provided with the force measuring device.
- 2) Prior to each reading, the dial gauge was lightly tapped on the center of the dial cover.
- 3) The reported uncertainty is at coverage factor  $k = 2$  which corresponds to a coverage probability of approximately 95% for a normal distribution, considering the relative error of different components such as Zero, Resolution, Repeatability, Interpolation and combining the uncertainty of applied force

  
Scientific Asst.  
(D. D. Magdum)  
Sr. Engineer  
(R.V. Tambad)  
Director  
(Dr. J.C. Padte)

**FRI**

Prosperity through research

# Fie Research Institute

22/44, Ganganagar P. O., ICHALKARANJI - 416 116.

(Dist.Kolhapur) Maharashtra State, INDIA.

■ Phone : (0230) 2441475 ■ Fax : (0230) 2440191

■ E-mail : fri@dataone.in ■ Website : www.fiegroupindia.com

**Accreditation No. NABL C0066**

## CALIBRATION CERTIFICATE OF FORCE MEASURING DEVICE

Date of calibration : 12/01/2014 Certificate No : FRI/01/14/6605  
Next calibration due on : 12/03/2016 Page No. : 2/2

Details of instrument under calibration Id No. 104 (KTPL/F/003)


Calibration Method The Dynamometer is calibrated in compression as per FRI Calibration procedure No. FRICAL/CAL/01 based on IS: 4169-1988 and results are tabulated below


Results The Calibration Results are valid for specific force steps/interpolation


Applied Force in kN	Dial Gauge Reading (Divisions)			Average
	Series1 at 0°	Series2 at 180°	Series3 at 360°	
5.000	50.8	50.8	50.8	50.8
10.000	102.6	102.6	102.6	102.6
15.000	154.2	154.4	154.4	154.3
20.000	206.4	206.4	206.6	206.5
25.000	258.6	258.6	258.8	258.7
30.000	310.8	310.8	311.0	310.9
35.000	362.8	362.8	363.0	362.9
40.000	414.8	414.8	414.8	414.8
45.000	466.2	466.2	466.4	466.3
50.000	517.8	518.0	518.0	517.9

**Classification :** The Classification of force proving instrument is as follows:

Class	Mode	From	To	Uncertainty of measurement
Class 0	Compression	50.000 kN	20.000 kN	± 0.080%
Class 1	Compression	50.000 kN	10.000 kN	± 0.127%
Class 2	Compression	50.000 kN	5.000 kN	± 0.235%

  
Scientific Asst.  
(D. D. Magdum)

  
Sr. Engineer  
(R.V. Tambad)

  
Director  
(Dr. J.C. Padte)